

Daily status, October 16

10/16/13

Summary:

Fixed Rlw.out.far pyrgometer once again.

Today will attempt to hardwire floor Zigbee motes (radiation and soil) to the DSM.

T/RH: RH.40m.near and RH.35m.rim perhaps a bit high.

P: ok

csat u,v: ok

csat ldiag: ok

csat w, tc: ok

kh2o: FLR variance high at night, but also tc'tc'

motes: FAR ok

NEAR ok

FLR all off and on; outages generally last 1:00 - 1:05 hrs.

Wetness: ok

radiation: Rpile.out.far ok

Tsoil: Tsoil.2.5cm.flr & far (linear avg) does not fit profile

Gsoil: ok

Qsoil: ok

Cvsoil: ok

2D sonic: ok