

# 19 May afternoon

We had 3 reasons to get out to some towers: fix 2m TRHs, take soil cores, and check on other coolers (since 2 have been found with water):

tse01: cooler had some water; took core

tnw02: 2m TRH fan wasn't working – replaced

tse02: cooler had lots of water; took core

tse04: 2m TRH fuse was blown – swapped from another port

tse06: 2m TRH swapped, but still not working. Likely a blown fuse, but is at 30m and didn't have climbing gear. Left new TRH in place.

tse07: took core

tnw09: cooler full of water; took core

tnw06: cooler full of water