

# licors set to 20sps

Both Licors just set to 20 samples/s using:

(Outputs (BW 20) (Delay 0) (RS232 (Freq 20.0)))