

Tower work - Kh20 cleaning. licor, par wetness partial

12July13 ~9:40-10:30

Cleaned Kryptons. Voltage have come back up, perhaps partly due to some drying

Cleaned Wetness, Par, Licors up tower

Wetness sensors had moisture sitting on them from dew or rains yesterday. High humidity could mean little evap.