

# Daily Report for Oct 10, 2013

Weather today is cold and partly cloudy. We've been having light snow off and on all day, but there has been no accumulation.

Power amp is at .143 mW

Nagios screen shows critical error with the profiler\_any\_profiler (checks 3 of 5 have failed). After checking again at 0500 LST, problem corrected itself and no flag appeared.

Profiler appears to be working and power spectral plots look good. Spoke with Bill and he suspects the error has to do possibly with some work that was done earlier. I'll keep an eye on it.

Nagios also show critical errors with the Lidar. System Tech arrived today and is working with Tim on replacing the computer. I imagine that is why the lidar is flagged red.

In looking the profiler display the dashboard buttons are all green, and the dwell counter is working properly.

An IOP has been called for tomorrow evening.

The generator is half full. I was going to refill, but couldn't figure out where how to hook up the tank to my rental car engine. Will ask Tim for help with this when I see him.

Internet connection is out again at main site, but just for the GAUS and ISS computers (same problem as we saw on Sunday night, cause unknown but data is all coming in.