

Service visit to NEAR

10/11/13, twh:

Arrived 4:39 pm. Cleaned kh2o and reseated all connectors (initially looked okay).

4:49 - 4:53 pm. Cleaned radiometers, wetness sensor.

4:58 - 5:08 pm. Replaced dessicant in pyrgeomters. The cylinders were screwed in so tight that I wonder whether they had ever been serviced.