

Daily Status 24 October

10/24/13

Summary:

- IOP last night, Dave says it was a good night
- cloudy this morning

Actions past 24 hours:

- SO and SC went to floor to fix P.sw.flr, battery replaced
- SO tried to improve RF comms for P.nne, refer to detail entry
- removed diode from soil battery system at near

To dos:

- SO looking into adding soil sensors to base

Sensor status:

T/RH: ok

P: ok

csat u,v: ok

csat ldiag: ok

csat w, tc: ok

kh2o: ok

motes: ok

Wetness: ok

radiation: ok

Tsoil: ok

Gsoil: ok

Qsoil: ok. Offsets based on yesterday's gravimetric samples make the data look reasonable (for the entire experiment).

Cvsoil: ok.

2D sonic: ok