

# Platteville Visit, data retrieval, Q7 swapped

4June09 ~10:15-11:15MDT

Pocketec Swapped. Brought unit taken back for data merge.

Photos taken. Sensor serial numbers confirmed.

Q7 swapped: We had installed the 'wrong' Rnet for the coefficients in the logger. Out=Q94197 In=Q99259