

Highland and West Valley Jan 1

Power cycled Highland at 7:00 AM. 4 inches of snow. Cleared off solar panels and radiometers.

Replaced batteries at West Valley at 8:15. 10 inches of snow. Cleared off solar panels and radiometers. ETI precip gauge is filled with snow and not melting into antifreeze. Needs to have known amount of antifreeze added. UU has antifreeze at Mtn Met Lab.