

Stn 3 Visit 24 Jan 2011

24Jan2011 14:55-15:55 Militzer/Horst

conditions: calm, upper 30's, high overcast, moisture coming in tonight

Purpose: Cleaning, Update mote software, check ETI

14:59-15:00 Rad. cleaned with methanol and distilled. Desiccant ok.

14:59-15:03 Mote software updated. Put fingernail polish on antenna connectors at interface of whitebox.

15:10-15:28 Mast lowered.

15:18 Kh₂O cleaned with methanol and distilled. Went from .92 to 2.0 volts.

Sonic wicks retracted a bit because they were extending a bit too far into the measurement paths.

15:39-15:58 A few adn/aup to try to communicate with ETI, but we were using the wrong approach to try it. Did not have the correct serial cable to tie-in directly to gauge. NOTE still need to confirm that the TO/1 temperature compensation is on.

Note: mast bolts ok, stakes pounded in on mast and rad stand