

Post cal TRH

Post cal results of temperature, temperature range 15C to -20C

Sensor	mean (error)	sd (error)	min to max
017	.019	.014	-.036 to .001
023	.017	.012	-.004 to .030
019	.016	.014	-.001 to .032
013	.000	.006	-.007 to .010
025	-.008	.011	-.034 to .002
024	.002	.008	-.011 to .013
009	-.035	.007	-.044 to -.022
028	-.018	.007	-.028 to -.004
027	-.011	.016	-.027 to .02

Post cal of RH, two runs from 10% to 90% at 10C and -4C, Gortex still on

Sensor	mean (error)	sd (Error)	min to max
017	-3.3	0.9	-4.8 to -1.4
023	-1.7	1.2	-3.8 to -0.4
019	-8.6	1.2	-11.1 to -6.3
013	0.1	0.5	-0.7 to 0.8
025	0.7	1.44	-2.2 to 2.2
024	1.1	1.7	-1.1 to 3.8
009	-0.6	0.6	-1.1 to 0.9
028	0.6	1.9	-3.6 to 2.6
027	1.2	1.6	-2.1 to 2.6

Location of SHT sensors through out the project. Some sensors replaced during the project were sent back to Boulder for repairs then the sensor was deployed again

DATE of CHANGE

STATION	11/28	12/07	12/12	12/13	1/15	11/16	1/17	1/18	1/27	1/28	2/6	2/7
1	13		9	26			21		13			13
2	15	21					24					24
3	20						19	23				23
4	19					25						25
5	18									9	17	17
6	23	24					27					27
7	17				28							28