

Tower Equipment Stripped

June 26, 2012

11:00am - 2:05pm MDT

Removal of all sensors on the Turbulence Tower today. Sonics, Licors, TRHs, pressure, four-component, batteries and dsm. Cables have remained on tower. Kept beacon on also.

Since we left cables on tower we had to mark cables on the dsm side so we know what instrument it was and which port. This was our organization codes using colored paper clips, zip ties and electrical tape. All levels have a sonic, TRH and Licor which are bundled together. This is the code for the bundle.

CSAT marked with zip tie

TRH marked with tape

Licor marked with nothing.

Heights are marked with paper clip colors.

45m - Green

30m - Pink

15m - Blue

8m - Yellow

2m - White

Additional random cables.

Ethernet - 1 SMALL zip tie, no clip

Rad22m - 2 SMALL zip ties, no clip

Pressure - 3 SMALL zip ties, no clip

dsm was taken down at 11:35am.

WE NEED PRESSURE TUBING!